



AlphaProTech N-95 Particulate Respirator

PFL® (Positive Facial Lock®) Design
with Magic Arch® and Twist Seal®



Donning Instructions



Step 1.
Grasp nosepiece at top of mask with one hand. Slide finger of other hand behind chin piece.



Step 2.
Open mask by pulling down and out on chin piece until mask is fully open. It is **IMPORTANT** that mask is fully open.



Step 3.
Place chin in pocket created. Stretch bands over back of head, positioning bottom band behind neck below ears and top band near crown of head.



Step 4.
Conform nosepiece to nose and face contours making sure Magic Arch® is fully extended away from nose and mouth.



Step 5.
Holding mask at nose-piece, place one finger inside the mask on the malleable aluminum chin piece, pull down gently so that mask is snug against face.



Step 6.
Pinch any excess material together under chin and fold over or twist.



Step 7.
Face Fit Check: To check fit, place both hands along the edges of the respirator and exhale. If air escapes around your nose, adjust the nose piece. If air leaks at the respirator edges, unfold or untwist the chin piece, repeat Steps 5 and 6.



Step 8.
PFL Mask correctly fit to face.

Warning

1. This product does not eliminate the risk of contracting any disease or infection. Change immediately if the respirator comes in contact with blood or fluids, including body fluids. OSHA has not set a permissible exposure level for airborne biohazards.

2. Follow all instructions and warnings on the use of this respirator. Failure to do so will reduce respirator effectiveness, wearer protection and may result in sickness or death. Wear during all times of exposure. For additional information refer to your supervisor and/or see dispenser box.

3. Before use, the wearer must be trained by the employer in proper respirator use in accordance with applicable safety and health standards for the contaminant and exposure level in the assigned work area.

Positive Face Lock, PFL, Magic Arch and Twist Seal are registered trademarks of AlphaProTech.

N-95 Surgical Respirator

Reorder Part No. 695

